

CERTIFICATE OF EO STERILE

COMPANY OF STERILIZATION: HAIMEN XIN CHUANG STERILIZATION SERVICE

Exporter: ZHEJIANG RENON MEDICAL INSTRUMENT CO., LTD

IMPORTER: Global Trade Technology - Comércio, Importação e Exportação Ltda - EPP.
Av. José Antonio Miziara n 221 - Distrito Ind. José Aparecido Tomé -
Jaboticabal - SP. - 14.874-002

Fone/Fax: (16) 4042-4000

CPNJ: 07.610.601/0001-65 INSCRIÇÃO ESTADUAL: 391.173.147.116

DATE OF STERILIZATION: 18TH APRIL 2024

GOODS	QTY	LOT	MFG date	Exp date
18278IND, INOCULATION LOOPS 10UL, ABS MATERIAL, EO STERILE, INDIVIDUAL PEEL PACK, 100 PCS/POLYBAG	10000packs	R202403	04/2024	04/2029
18276IND, INOCULATION LOOPS 1UL, ABS MATERIAL, EO STERILE, INDIVIDUAL PEEL PACK, 100 PCS/POLYBAG	10000packs	R202403	04/2024	04/2029

EO DOZE:18KG

INVOICE NO: RN2024-024

We hereby certificate that the above goods have been sterilized by EO, with the lowest temperature of 45°C, highest temperature of 60°C sterilization duration of 12 hours

2024-4-18



CALIBRATION CERTIFICATE

ACCORDING TO Q/320684FHN13-2007

Commodity: LOOP 1UL AND LOOP 10UL

MATERIAL:ABS

THE LOOP HEAD SIZE accord with standard GB/T14486

IUL HEAD DIAMETER 2MM WHICH CAN TAKE IUL LIQUID

10UL HEAD DIAMTER 5.8MM WHICH CAN TAKE 1OUL LIQUID

LOOP HEAD :BAR NOT ALLOWED.

2 BULB: BIG BULB NOT ALLOWED

3 COLOR: NO COLOR distinctness

STERILE BY GAMMA RAY ACCORDING TO GB 15980

ACCORDING TO THE STANDARD:

THE GOODS PASS THE TEST INVOICE NO: RN2024-024

LOT NO: Lot R202403

QTY: 2000PACKS

Date 2024-4-26

浙江理能医疗器械有限公司
ZHEJIANG RENON MEDICAL INSTRUMENT CO.,LTD



CERTIFICATE OF EO STERILE

COMPANY OF STERILIZATION: HAIMEN XIN CHUANG STERILIZATION SERVICE

Exporter: ZHEJIANG RENON MEDICAL INSTRUMENT CO.,LTD.

IMPORTER: Global Trade Technology - Comércio, Importação e Exportação Ltda - EPP.

Av. José Antonio Miziara n 221 - Distrito Ind. José Aparecido Tomé -

Jaboticabal - SP. - 14.874-002

Fone/Fax: (16) 4042-4000

CPNJ: 07.610.601/0001-65 INSCRIÇÃO ESTADUAL: 391.173.147.116

DATE OF STERILIZATION: 25TH AUGUST 2024

GOODS	QTY	LOT	MFG date	Exp date
18278IND INOCULATION LOOPS 10UL, ABS MATERIAL, EO STERILE, INDIVIDUAL PEEL PACK, 100 PCS/POLYBAG	3000packs	R202408	08/2024	08/2029
18276IND INOCULATION LOOPS 1UL, ABS MATERIAL, EO STERILE, INDIVIDUAL PEEL PACK, 100 PCS/POLYBAG	3000packs	R202408	08/2024	08/2029

EO DOZE:18KG

INVOICE NO: RN2024-149

We hereby certificate that the above goods have been sterilized by EO, with the lowest temperature of 45°C, highest temperature of 60°C sterilization duration of 12 hours

2024-8-25



郑建达

CALIBRATION CERTIFICATE

ACCORDING TO Q/320684FHN13-2007

Commodity: LOOP 1UL AND LOOP 10UL

MATERIAL:ABS

THE LOOP HEAD SIZE accord with standard GB/T14486

IUL HEAD DIAMETER 2MM WHICH CAN TAKE IUL LIQUID

10UL HEAD DIAMTER 5.8MM WHICH CAN TAKE 10UL LIQUID

LOOP HEAD :BAR NOT ALLOWED.

2 BULB: BIG BULB NOT ALLOWED

3 COLOR: NO COLOR distinctness

STERILE BY GAMMA RAY ACCORDING TO GB 15980

ACCORDING TO THE STANDARD:

THE GOODS PASS THE TEST INVOICE NO: RN2024-149

LOT NO: Lot R202408

QTY: 1UL: 3000PACKS

10UL: 3000PACKS

Date 2024-08-25

浙江理能医疗器械有限公司
ZHEJIANG RENON MEDICAL INSTRUMENT CO.,LTD

