

NANTONG MICHAEL IRRADIATION CO., LTD


CERTIFICATE OF GAMMA STERILIZATION

Date:2024-8-15

STERILE COMPANY	NANTONG MICHAEL IRRADIATION CO., LTD
EXPORTER	ZHEJIANG RENON MEDICAL INSTRUMENT CO., LTD
IMPORTER	GLOBAL TRADE TECHNOLOGY COMERCIO IMPORTACO E EXPORTACAO LTDA AVENIDA DOUTOR JOSÉ ANTÔNIO MIZIARA, 221 - DISTRITO INDUSTRIAL JABOTICABAL SP - CITY ZIP CODE 14874-002
灭菌日期 DATE OF STERILIZATION	14 ^h AUGUST 2024
Invoice Number: RN2024-123	
Product: GT-GP1000LR: pipette tip 1000ul long in rack - Type Gilson - 01 RACK 96UN	
QTY:500UN lot:20240708, mfg date: 07/2024 exp date:07/2026	
GT-GP15: Centrifuge tube 15ml, DNA/RNA free QTY:300PACKS lot:202408, mfg date: 08/2024 exp date:08/2026	
要求 GAMMA 剂量 MIN GAMMA DOSE 25KGY MAXIUM GAMMA DOSE:40KGY	
南通市迈克尔辐照有限公司 特此证明上述产品已经本公司 GAMMA 射线消毒 NANTONG MICHAEL IRRADIATION CO., LTD hereby warranty that all products mentioned above have been sterilized by GAMMA RAY	

STAMP:

SIGNATURE



陈利康


NANTONG MICHAEL IRRADIATION CO., LTD

CERTIFICATE OF GAMMA STERILIZATION

Date:2024-9-15

STERILE COMPANY	NANTONG MICHAEL IRRADIATION CO., LTD
EXPORTER	ZHEJIANG RENON MEDICAL INSTRUMENT CO., LTD
IMPORTER	GLOBAL TRADE TECHNOLOGY COMERCIO IMPORTACO E EXPORTACAO LTDA AVENIDA DOUTOR JOSÉ ANTÔNIO MIZIARA, 221 - DISTRITO INDUSTRIAL JABOTICABAL SP - CITY ZIP CODE 14874-002
灭菌日期 DATE OF STERILIZATION	14 ^h SEPTEMBER 2024
Invoice Number: RN2024-165 Product: GT-GP15: Centrifuge tube 15ml, DNA/RNA free QTY:700Pct lot:202409, mfg date: 09/2024 exp date:09/2026	
要求 GAMMA 剂量 MIN GAMMA DOSE 25KGY MAXIUM GAMMA DOSE:40KGY	
南通市迈克尔辐照有限公司 特此证明上述产品已经本公司 GAMMA 射线消毒 NANTONG MICHAEL IRRADIATION CO., LTD hereby warranty that all products mentioned above have been sterilized by GAMMA RAY	

STAMP:



SIGNATURE

陈利康